

General Care

Finger Pulse Oximeter PI F02T

Product specification

- Anti-electric Shock Type: Internally powered equipment
- Anti-electric Shock Degree: Type BF equipment

Measure Accuracy:	Hemoglobin Saturation	2% (80% - 100%) 3% (70% - 80%) Unspecified (≤70%)
	Pulse rate	25bpm ~ 250bpm ± 2bpm

Kemenkes RI AKL 20502918057