www.AUTOCHECKco.id

Multi-Monitoring System

AUTOCHECK®

STRIP

Auto Strip Recognition CHOL, UA

- No Coding for Blood Glucose Strip
- Total Cholesterol 26 sec.
- Blood Glucose 5 sec.
- Uric Acid 15 sec.



Self-testing for in-vitro diagnostic use only.
Please read your user's manual before using this product.

CE 0197

KEMENKES RI AKL 20101810112

KEMENKES RI AKL 20101311323

KEMENKES RI AKL 20101810023

SPECIFICATION

Total Cholesterol Test Strip

Principle	Electrochemical Biosensor
Test Sample	Fingertip Capillary Whole Blood
Measuring Time	26 Seconds
Measuring Range	100 - 400 mg/dL (2.59 - 10.35 mmol/L)
Sample Volume	1 µL
Memory	50 Test Result
Content	15 + 10 Strips
Kemenkes RI AKL	20101810112

Blood Glucose Test Strip

Principle	Electrochemical Biosensor
Test Sample	Fingertip, Palm, and Forearm Capillary Whole Blood
Measuring Time	5 Seconds
Measuring Range	20 - 600 mg/dL (1.1 - 33.3 mmol/L)
Sample Volume	0.5 µL
Memory	360 Test Result
Content	25 + 25 Strips
Kemenkes RI AKL	20101311323

Uric Acid Test Strip

Principle	Electrochemical Biosensor
Test Sample	Fingertip Capillary Whole Blood
Measuring Time	15 Seconds
Measuring Range	3 - 20 mg/dL (0.18 - 1.19 mmol/L)
Sample Volume	1 µL
Memory	50 Test Result
Content	25 + 25 Strips
Kemenkes RI AKL	20101810023

Product Warranty

Strip Unopened Shelf Life	*(from Manufacturing Day)*
- Total Cholesterol	16 Months
- Blood Glucose	24 Months
- Uric Acid	18 Months
Strip Opened Shelf Life	
- Total Cholesterol	3 Months
- Blood Glucose	6 Months
- Uric Acid	3 Months

GEA medical
PT. MEGA PRATAMA MEDICALINDO
Jakarta - INDONESIA